

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050422\
 Data File : VN072325.D
 Acq On : 04 May 2022 14:02
 Operator : JC\MD
 Sample : N2680-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B4-GW

Quant Time: May 05 02:19:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	230570	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	449853	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	491929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	188513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	185418	52.381	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.760%	
35) Dibromofluoromethane	8.022	113	146050	50.724	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.440%	
50) Toluene-d8	10.439	98	582697	51.502	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.000%	
62) 4-Bromofluorobenzene	12.727	95	226626	60.348	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 120.700%	
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	3864	2.532	ug/l #	71
64) Tetrachloroethene	10.980	164	1372	0.463	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050422\
Data File : VN072325.D
Acq On : 04 May 2022 14:02
Operator : JC\MD
Sample : N2680-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B4-GW

Quant Time: May 05 02:19:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

