

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050422\
 Data File : VN072326.D
 Acq On : 04 May 2022 14:26
 Operator : JC\MD
 Sample : N2672-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB

Quant Time: May 05 02:19:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	230586	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	446871	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	476613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	181079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	190794	53.896	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 107.800%		
35) Dibromofluoromethane	8.022	113	142336	49.764	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 99.520%		
50) Toluene-d8	10.439	98	577293	51.365	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 102.740%		
62) 4-Bromofluorobenzene	12.727	95	217552	58.318	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 116.640%		
Target Compounds						Qvalue
16) Acetone	4.298	43	9478	6.210	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050422\
Data File : VN072326.D
Acq On : 04 May 2022 14:26
Operator : JC\MD
Sample : N2672-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB

Quant Time: May 05 02:19:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

