

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050422\
 Data File : VN072332.D
 Acq On : 04 May 2022 17:01
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 05 02:21:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	228845	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	446055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	474505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	182514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	181676	51.711	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.420%	
35) Dibromofluoromethane	8.016	113	139603	48.898	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.800%	
50) Toluene-d8	10.439	98	570869	50.887	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.780%	
62) 4-Bromofluorobenzene	12.727	95	217550	58.424	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 116.840%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050422\
Data File : VN072332.D
Acq On : 04 May 2022 17:01
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 05 02:21:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

