

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050423\
 Data File : VN077587.D
 Acq On : 04 May 2023 18:38
 Operator : JC\MD
 Sample : 02566-05DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1989DL

Quant Time: May 05 04:12:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	757155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1403540	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1299156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	544201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.578	65	445992	45.141	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.280%		
35) Dibromofluoromethane	8.172	113	420694	49.870	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.740%		
50) Toluene-d8	10.566	98	1683874	52.082	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.160%		
62) 4-Bromofluorobenzene	12.854	95	619146	58.532	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	117.060%		
Target Compounds					Qvalue	
16) Acetone	4.449	43	27633	8.791	ug/l	97
20) Methylene Chloride	5.284	84	12515	1.256	ug/l #	84
40) Benzene	8.613	78	1138657	30.669	ug/l	100
43) Isopropyl Acetate	8.701	43	39057	2.438	ug/l	93
52) Toluene	10.630	92	629463	25.880	ug/l	100
67) Ethyl Benzene	11.966	91	93212	1.926	ug/l	97
68) m/p-Xylenes	12.072	106	207941	11.055	ug/l	96
69) o-Xylene	12.401	106	214597	11.599	ug/l	97
80) 1,3,5-Trimethylbenzene	13.177	105	60446	1.612	ug/l	97
84) 1,2,4-Trimethylbenzene	13.489	105	218297	6.233	ug/l	88
95) Naphthalene	15.642	128	1322029	53.914	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050423\
Data File : VN077587.D
Acq On : 04 May 2023 18:38
Operator : JC\MD
Sample : 02566-05DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1989DL

Quant Time: May 05 04:12:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

