

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050520\
 Data File : VN061292.D
 Acq On : 5 May 2020 15:18
 Operator : JC/MD
 Sample : L2462-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AOC-14

Quant Time: May 06 05:23:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	102340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	184730	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	174776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	76628	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	83638	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
35) Dibromofluoromethane	7.55	113	54202	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
50) Toluene-d8	10.06	98	226884	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	12.38	95	85171	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
Target Compounds						
16) Acetone	3.77	43	6279	4.603	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050520\
Data File : VN061292.D
Acq On : 5 May 2020 15:18
Operator : JC/MD
Sample : L2462-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AOC-14

Quant Time: May 06 05:23:29 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

