

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050520\
 Data File : VN061296.D
 Acq On : 5 May 2020 16:51
 Operator : JC/MD
 Sample : L2514-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: May 06 05:54:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	97656	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	177610	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	165218	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	74500	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	80989	52.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.82%	
35) Dibromofluoromethane	7.55	113	52917	50.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.78%	
50) Toluene-d8	10.06	98	217669	50.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.68%	
62) 4-Bromofluorobenzene	12.38	95	80988	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
Target Compounds						
11) Tert butyl alcohol	4.73	59	31687	153.082	ug/l	99
16) Acetone	3.78	43	8123	8.250	ug/l	94
19) Methyl tert-butyl Ether	4.99	73	20767	5.275	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050520\
Data File : VN061296.D
Acq On : 5 May 2020 16:51
Operator : JC/MD
Sample : L2514-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: May 06 05:54:55 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

