

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
 Data File : VN066987.D
 Acq On : 5 May 2021 14:32
 Operator : JC/MD
 Sample : M2253-05 400X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41764

Manual Integrations
 APPROVED

MMDadoda
 5/6/2021 5:40:19 PM

Quant Time: May 06 04:49:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	788755	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.517	114	1338581	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	1102650	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	422589	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	572707	46.13	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.26%		
35) Dibromofluoromethane	7.514	113	421969	46.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.84%		
50) Toluene-d8	10.030	98	1669334	48.22	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.44%		
62) 4-Bromofluorobenzene	12.350	95	520347m	44.39	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	88.78%		
Target Compounds						Qvalue
16) Acetone	3.760	43	10051m	2.30	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
Data File : VN066987.D
Acq On : 5 May 2021 14:32
Operator : JC/MD
Sample : M2253-05 400X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41764

Manual Integrations
APPROVED

MMDadoda
5/6/2021 5:40:19 PM

Quant Time: May 06 04:49:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

