

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
 Data File : VN066989.D
 Acq On : 5 May 2021 15:22
 Operator : JC/MD
 Sample : PB136154ZHE#17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136154ZHE#17

Manual Integrations
 APPROVED

MMDadoda
 5/6/2021 5:40:20 PM

Quant Time: May 06 04:50:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	815974	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	1406759	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	1286639	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	442773	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.957	65	597117	46.49	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.98%		
35) Dibromofluoromethane	7.514	113	437970	45.85	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.70%		
50) Toluene-d8	10.030	98	1814199	49.87	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.74%		
62) 4-Bromofluorobenzene	12.347	95	572005	46.44	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	92.88%		
Target Compounds						Qvalue
16) Acetone	3.760	43	236852	52.42	ug/l	97
18) Methyl Acetate	4.272	43	16735m	1.29	ug/l	
20) Methylene Chloride	4.470	84	452012	39.44	ug/l	90
43) Isopropyl Acetate	8.099	43	101178	4.17	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
Data File : VN066989.D
Acq On : 5 May 2021 15:22
Operator : JC/MD
Sample : PB136154ZHE#17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136154ZHE#17

Manual Integrations
APPROVED

MMDadoda
5/6/2021 5:40:20 PM

Quant Time: May 06 04:50:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

