

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
 Data File : VN066990.D
 Acq On : 5 May 2021 15:46
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 06 04:50:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.589	168	759985	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.517	114	1296015	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	1085507	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	420582	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	549131	45.90	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.80%
35) Dibromofluoromethane	7.511	113	412825	46.91	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.82%
50) Toluene-d8	10.030	98	1635396	48.80	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.60%
62) 4-Bromofluorobenzene	12.350	95	429652	37.86	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	75.72%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
Data File : VN066990.D
Acq On : 5 May 2021 15:46
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 06 04:50:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

