

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
 Data File : VN066991.D
 Acq On : 5 May 2021 16:44
 Operator : JC/MD
 Sample : PB136154ZHE#18
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136154ZHE#18

Quant Time: May 06 04:50:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	716984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.517	114	1245633	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	1135712	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	463254	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	592344	52.49	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.98%			
35) Dibromofluoromethane	7.512	113	435249	51.46	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.92%			
50) Toluene-d8	10.030	98	1748701	54.29	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.58%			
62) 4-Bromofluorobenzene	12.350	95	581097	53.28	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 106.56%			
Target Compounds					Qvalue	
16) Acetone	3.757	43	181066	45.61	ug/l	98
20) Methylene Chloride	4.473	84	349854	34.74	ug/l	91
43) Isopropyl Acetate	8.096	43	74047	3.45	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
Data File : VN066991.D
Acq On : 5 May 2021 16:44
Operator : JC/MD
Sample : PB136154ZHE#18
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136154ZHE#18

Quant Time: May 06 04:50:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

