

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
 Data File : VN066992.D
 Acq On : 5 May 2021 17:05
 Operator : JC/MD
 Sample : PB136154ZHE#19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136154ZHE#19

Quant Time: May 06 04:51:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	598580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	1025330	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	950000	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	344466	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	488494	51.85	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.70%			
35) Dibromofluoromethane	7.514	113	354302	50.89	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.78%			
50) Toluene-d8	10.030	98	1419414	53.53	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.06%			
62) 4-Bromofluorobenzene	12.350	95	450630	50.19	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 100.38%			
Target Compounds					Qvalue	
16) Acetone	3.757	43	167334	50.49	ug/l	95
18) Methyl Acetate	4.267	43	11172	1.18	ug/l #	88
20) Methylene Chloride	4.476	84	320059	38.07	ug/l	92
43) Isopropyl Acetate	8.099	43	70941	4.01	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
Data File : VN066992.D
Acq On : 5 May 2021 17:05
Operator : JC/MD
Sample : PB136154ZHE#19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136154ZHE#19

Quant Time: May 06 04:51:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

