

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
 Data File : VN066993.D
 Acq On : 5 May 2021 17:27
 Operator : JC/MD
 Sample : PB136154ZHE#20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136154ZHE#20

Quant Time: May 06 04:51:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.595	168	558932	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	939732	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	863237	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	294091	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	449477	51.09	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.18%	
35) Dibromofluoromethane	7.517	113	324040	50.78	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.56%	
50) Toluene-d8	10.030	98	1321138	54.36	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 108.72%	
62) 4-Bromofluorobenzene	12.350	95	409374	49.75	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.50%	
Target Compounds						
						Qvalue
16) Acetone	3.762	43	152076	49.14	ug/l	98
18) Methyl Acetate	4.267	43	10614	1.20	ug/l #	82
20) Methylene Chloride	4.476	84	290447	37.00	ug/l	91
43) Isopropyl Acetate	8.096	43	61953	3.83	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
Data File : VN066993.D
Acq On : 5 May 2021 17:27
Operator : JC/MD
Sample : PB136154ZHE#20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136154ZHE#20

Quant Time: May 06 04:51:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

