

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
 Data File : VN066997.D
 Acq On : 5 May 2021 19:03
 Operator : JC/MD
 Sample : M2254-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: May 06 04:52:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	472323	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	806346	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	761427	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	300196	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	399600	53.75	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.50%	
35) Dibromofluoromethane	7.514	113	282849	51.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.32%	
50) Toluene-d8	10.030	98	1138156	54.58	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 109.16%	
62) 4-Bromofluorobenzene	12.347	95	359529	50.92	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.84%	
Target Compounds						
						Qvalue
16) Acetone	3.757	43	62257	23.80	ug/l	95
18) Methyl Acetate	4.261	43	10612	1.42	ug/l #	78
20) Methylene Chloride	4.476	84	61237	9.23	ug/l	94
43) Isopropyl Acetate	8.096	43	57054	4.11	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050521\
Data File : VN066997.D
Acq On : 5 May 2021 19:03
Operator : JC/MD
Sample : M2254-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: May 06 04:52:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

