

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050522\
 Data File : VN072336.D
 Acq On : 05 May 2022 11:09
 Operator : JC\MD
 Sample : VN0505MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0505MBL01

Quant Time: May 06 01:33:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	221592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	430574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	459643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	174893	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	182676	53.698	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.400%
35) Dibromofluoromethane	8.016	113	138996	50.436	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.880%
50) Toluene-d8	10.439	98	553615	51.123	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.240%
62) 4-Bromofluorobenzene	12.727	95	207929	57.848	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	115.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050522\
Data File : VN072336.D
Acq On : 05 May 2022 11:09
Operator : JC\MD
Sample : VN0505MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0505MBL01

Quant Time: May 06 01:33:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

