

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050522\
 Data File : VN072341.D
 Acq On : 05 May 2022 13:41
 Operator : JC\MD
 Sample : N2641-04
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-24

Quant Time: May 06 01:35:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	216892	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	411655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	449436	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	190565	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	168078	50.477	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.960%	
35) Dibromofluoromethane	8.016	113	129132	49.010	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.020%	
50) Toluene-d8	10.439	98	527802	50.979	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.960%	
62) 4-Bromofluorobenzene	12.727	95	215822	62.804	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 125.600%	
Target Compounds						
25) 2-Butanone	7.351	43	29851	13.056	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050522\
Data File : VN072341.D
Acq On : 05 May 2022 13:41
Operator : JC\MD
Sample : N2641-04
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-24

Quant Time: May 06 01:35:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

