

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050522\
 Data File : VN072342.D
 Acq On : 05 May 2022 14:05
 Operator : JC\MD
 Sample : N2619-06ME
 Misc : 4.66g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB03-13-14-FTME

Quant Time: May 06 01:35:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.075	168	237862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	458314	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	495030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	220649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	183905	50.361	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 100.720%		
35) Dibromofluoromethane	8.016	113	141668	48.294	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 96.580%		
50) Toluene-d8	10.439	98	582341	50.521	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 101.040%		
62) 4-Bromofluorobenzene	12.727	95	244231	63.835	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 127.680%		
Target Compounds						Qvalue
25) 2-Butanone	7.357	43	29492	11.762	ug/l #	87
31) Cyclohexane	8.069	56	7400	1.517	ug/l #	55
67) Ethyl Benzene	11.839	91	90964	5.215	ug/l	98
78) n-propylbenzene	12.916	91	42069	1.960	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050522\
Data File : VN072342.D
Acq On : 05 May 2022 14:05
Operator : JC\MD
Sample : N2619-06ME
Misc : 4.66g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB03-13-14-FTME

Quant Time: May 06 01:35:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

