



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050523\
 Data File : VN077589.D
 Acq On : 05 May 2023 13:08
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042823W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Fri Apr 28 16:20:59 2023



AutoFind: Scans 1855, 1856, 1857; Background Corrected with Scan 1846

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	27027	PASS
75	95	30	60	55.0	84549	PASS
95	95	100	100	100.0	153856	PASS
96	95	5	9	6.9	10676	PASS
173	174	0.00	2	1.1	1177	PASS
174	95	50	100	67.1	103285	PASS
175	174	5	9	7.5	7701	PASS
176	174	95	101	98.2	101405	PASS
177	176	5	9	6.4	6526	PASS