

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN050525\  
 Data File : VN086480.D  
 Acq On : 05 May 2025 12:38  
 Operator : JC\MD  
 Sample : VN0505WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0505WBL01

Quant Time: May 06 01:43:53 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N041525W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Apr 16 04:19:23 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 8.224  | 168   | 156438   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.100  | 114   | 296282   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.865 | 117   | 275019   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.788 | 152   | 110905   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 8.577  | 65    | 111180   | 49.011   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 98.020%  |
| 35) Dibromofluoromethane    | 8.165  | 113   | 82766    | 60.190   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 120.380% |
| 50) Toluene-d8              | 10.565 | 98    | 373128   | 50.772   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 101.540% |
| 62) 4-Bromofluorobenzene    | 12.847 | 95    | 126773   | 47.294   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =     | 94.580%  |

Target Compounds Qvalue

-----  
 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN050525\  
Data File : VN086480.D  
Acq On : 05 May 2025 12:38  
Operator : JC\MD  
Sample : VN0505WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0505WBL01

Quant Time: May 06 01:43:53 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N041525W.M  
Quant Title : SW846 8260  
QLast Update : Wed Apr 16 04:19:23 2025  
Response via : Initial Calibration

