

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050525\
 Data File : VN086478.D
 Acq On : 05 May 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 06 01:43:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	69	0.00
2 T	Dichlorodifluoromethane	0.593	0.536	9.6	63	0.00
3 P	Chloromethane	0.861	0.805	6.5	74	0.00
4 C	Vinyl Chloride	0.822	0.759	7.7#	66	0.00
5 T	Bromomethane	0.370	0.461	-24.6	90	0.00
6 T	Chloroethane	0.550	0.523	4.9	72	0.00
7 T	Trichlorofluoromethane	0.911	0.869	4.6	70	0.00
8 T	Diethyl Ether	0.398	0.421	-5.8	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.551	0.537	2.5	71	0.00
10 T	Methyl Iodide	0.606	0.738	-21.8	85	0.00
11 T	Tert butyl alcohol	0.132	0.140	-6.1	78	0.00
12 CM	1,1-Dichloroethene	0.588	0.557	5.3#	70	0.00
13 T	Acrolein	0.075	0.089	-18.7	101	0.00
14 T	Allyl chloride	1.046	0.919	12.1	68	0.00
15 T	Acrylonitrile	0.327	0.351	-7.3	76	0.00
16 T	Acetone	0.290	0.278	4.1	77	0.00
17 T	Carbon Disulfide	1.762	1.533	13.0	68	0.00
18 T	Methyl Acetate	0.870	0.830	4.6	73	0.00
19 T	Methyl tert-butyl Ether	2.163	2.360	-9.1	81	0.00
20 T	Methylene Chloride	0.670	0.729	-8.8	81	0.00
21 T	trans-1,2-Dichloroethene	0.615	0.632	-2.8	76	0.00
22 T	Diisopropyl ether	2.276	2.355	-3.5	77	0.00
23 T	Vinyl Acetate	1.591	1.722	-8.2	80	0.00
24 P	1,1-Dichloroethane	1.201	1.259	-4.8	77	0.00
25 T	2-Butanone	0.451	0.469	-4.0	76	0.00
26 T	2,2-Dichloropropane	1.075	1.069	0.6	75	0.00
27 T	cis-1,2-Dichloroethene	0.764	0.829	-8.5	80	0.00
28 T	Bromochloromethane	0.509	0.533	-4.7	74	0.00
29 T	Tetrahydrofuran	0.302	0.311	-3.0	77	0.00
30 C	Chloroform	1.180	1.306	-10.7#	82	0.00
31 T	Cyclohexane	1.175	0.999	15.0	65	0.00
32 T	1,1,1-Trichloroethane	1.009	1.049	-4.0	76	0.00
33 S	1,2-Dichloroethane-d4	0.725	0.817	-12.7	78	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	72	0.00
35 S	Dibromofluoromethane	0.232	0.264	-13.8	88	0.00
36 T	1,1-Dichloropropene	0.463	0.446	3.7	72	0.00
37 T	Ethyl Acetate	0.488	0.487	0.2	78	0.00
38 T	Carbon Tetrachloride	0.441	0.442	-0.2	75	0.00
39 T	Methylcyclohexane	0.551	0.492	10.7	67	0.00
40 TM	Benzene	1.485	1.541	-3.8	79	0.00
41 T	Methacrylonitrile	0.283	0.294	-3.9	80	0.00
42 TM	1,2-Dichloroethane	0.474	0.533	-12.4	86	0.00
43 T	Isopropyl Acetate	1.177	0.947	19.5	79	0.00
44 TM	Trichloroethene	0.352	0.372	-5.7	79	0.00
45 C	1,2-Dichloropropane	0.364	0.382	-4.9#	79	0.00
46 T	Dibromomethane	0.232	0.271	-16.8	86	0.00
47 T	Bromodichloromethane	0.500	0.559	-11.8	84	0.00
48 T	Methyl methacrylate	0.429	0.426	0.7	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050525\
 Data File : VN086478.D
 Acq On : 05 May 2025 11:32
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 06 01:43:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	87	0.00
50 S	Toluene-d8	1.240	1.280	-3.2	73	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.527	-5.4	78	0.00
52 CM	Toluene	0.927	0.979	-5.6#	79	0.00
53 T	t-1,3-Dichloropropene	0.558	0.627	-12.4	82	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.648	-5.7	81	0.00
55 T	1,1,2-Trichloroethane	0.333	0.382	-14.7	86	0.00
56 T	Ethyl methacrylate	0.613	0.661	-7.8	82	0.00
57 T	1,3-Dichloropropane	0.592	0.663	-12.0	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.291	0.277	4.8	69	0.00
59 T	2-Hexanone	0.371	0.388	-4.6	78	0.00
60 T	Dibromochloromethane	0.366	0.426	-16.4	87	0.00
61 T	1,2-Dibromoethane	0.336	0.394	-17.3	87	0.00
62 S	4-Bromofluorobenzene	0.452	0.478	-5.8	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	75	0.00
64 T	Tetrachloroethene	0.385	0.386	-0.3	78	0.00
65 PM	Chlorobenzene	1.116	1.173	-5.1	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.402	-9.2	87	0.00
67 C	Ethyl Benzene	2.014	1.999	0.7#	78	0.00
68 T	m/p-Xylenes	0.754	0.772	-2.4	80	0.00
69 T	o-Xylene	0.746	0.767	-2.8	81	0.00
70 T	Styrene	1.243	1.334	-7.3	82	0.00
71 P	Bromoform	0.277	0.312	-12.6	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
73 T	Isopropylbenzene	4.090	3.762	8.0	78	0.00
74 T	N-amyl acetate	2.036	1.853	9.0	80	0.00
75 P	1,1,2,2-Tetrachloroethane	1.176	1.187	-0.9	88	0.00
76 T	1,2,3-Trichloropropane	1.180	1.136	3.7	84	0.00
77 T	Bromobenzene	0.906	0.962	-6.2	87	0.00
78 T	n-propylbenzene	4.820	4.463	7.4	77	0.00
79 T	2-Chlorotoluene	3.016	2.903	3.7	81	0.00
80 T	1,3,5-Trimethylbenzene	3.348	3.185	4.9	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.491	0.514	-4.7	86	0.00
82 T	4-Chlorotoluene	2.996	2.949	1.6	81	0.00
83 T	tert-Butylbenzene	2.901	2.643	8.9	78	0.00
84 T	1,2,4-Trimethylbenzene	3.408	3.330	2.3	81	0.00
85 T	sec-Butylbenzene	4.023	3.690	8.3	76	0.00
86 T	p-Isopropyltoluene	3.316	3.176	4.2	78	0.00
87 T	1,3-Dichlorobenzene	1.704	1.787	-4.9	86	0.00
88 T	1,4-Dichlorobenzene	1.716	1.810	-5.5	86	0.00
89 T	n-Butylbenzene	2.939	2.742	6.7	74	0.00
90 T	Hexachloroethane	0.558	0.522	6.5	78	0.00
91 T	1,2-Dichlorobenzene	1.652	1.755	-6.2	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.237	0.233	1.7	81	0.00
93 T	1,2,4-Trichlorobenzene	0.791	0.848	-7.2	85	0.00
94 T	Hexachlorobutadiene	0.299	0.276	7.7	76	0.00
95 T	Naphthalene	2.804	2.877	-2.6	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050525\
Data File : VN086478.D
Acq On : 05 May 2025 11:32
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 06 01:43:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.800	-6.8	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6