

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050619\
 Data File : VN055406.D
 Acq On : 6 May 2019 9:54
 Operator : JC/SP
 Sample : VN0506MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0506MBL01

Quant Time: May 07 01:23:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	513149	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	897365	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	827242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	232435	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	355774	53.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.00%	
35) Dibromofluoromethane	7.59	113	293905	49.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.80%	
50) Toluene-d8	10.09	98	1115193	52.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.86%	
62) 4-Bromofluorobenzene	12.40	95	357626	53.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.60%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050619\
Data File : VN055406.D
Acq On : 6 May 2019 9:54
Operator : JC/SP
Sample : VN0506MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0506MBL01

Quant Time: May 07 01:23:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

