

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050619\
 Data File : VN055415.D
 Acq On : 6 May 2019 13:39
 Operator : JC/SP
 Sample : K2658-04 10X
 Misc : 6.25µ/5mL/100µL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-SB-E6(17-20)

Quant Time: May 07 02:07:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	500587	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	901672	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	846545	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	344450	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	352304	53.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.60%	
35) Dibromofluoromethane	7.59	113	281281	47.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.06%	
50) Toluene-d8	10.09	98	1117358	52.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.58%	
62) 4-Bromofluorobenzene	12.40	95	479082	71.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	143.46%	
Target Compounds						
73) Isopropylbenzene	12.25	105	728735	25.008	ug/l	100
78) n-propylbenzene	12.59	91	2161923	61.102	ug/l	99
83) tert-Butylbenzene	12.99	119	149627	5.222	ug/l	94
85) sec-Butylbenzene	13.17	105	1125626	38.351	ug/l #	66
89) n-Butylbenzene	13.61	91	555210	32.389	ug/l	90
95) Naphthalene	15.13	128	7368	4.047	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050619\
Data File : VN055415.D
Acq On : 6 May 2019 13:39
Operator : JC/SP
Sample : K2658-04 10X
Misc : 6.25g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-E6(17-20)

Quant Time: May 07 02:07:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

