

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050619\
 Data File : VN055418.D
 Acq On : 6 May 2019 16:14
 Operator : JC/SP
 Sample : K2639-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-08-050319-A

Quant Time: May 07 02:15:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	528810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	939821	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	883517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	297199	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	375120	54.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.44%	
35) Dibromofluoromethane	7.59	113	301473	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
50) Toluene-d8	10.09	98	1160025	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
62) 4-Bromofluorobenzene	12.40	95	402487	57.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.64%	
Target Compounds						
16) Acetone	3.83	43	14528	6.740	ug/l #	87
20) Methylene Chloride	4.55	84	14862	2.259	ug/l	96
43) Isopropyl Acetate	8.17	43	12856	1.159	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050619\
Data File : VN055418.D
Acq On : 6 May 2019 16:14
Operator : JC/SP
Sample : K2639-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-08-050319-A

Quant Time: May 07 02:15:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

