

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050619\
 Data File : VN055422.D
 Acq On : 6 May 2019 17:54
 Operator : JC/SP
 Sample : K2680-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SSSI-0502-S-TCLP-8

Quant Time: May 07 02:25:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	492259	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	771442	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	621145	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	179822	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	315615	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
35) Dibromofluoromethane	7.59	113	246432	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
50) Toluene-d8	10.09	98	909739	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
62) 4-Bromofluorobenzene	12.40	95	265982	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
Target Compounds						
16) Acetone	3.83	43	7621	3.798	ug/l	99
20) Methylene Chloride	4.55	84	707785	115.562	ug/l	98
43) Isopropyl Acetate	8.17	43	10670	1.172	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050619\
Data File : VN055422.D
Acq On : 6 May 2019 17:54
Operator : JC/SP
Sample : K2680-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SSSI-0502-S-TCLP-8

Quant Time: May 07 02:25:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

