


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN050619\
Data File : VN055404.D
Acq On : 6 May 2019 8:23
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Title : SW846 8260
Last Update : Tue Apr 30 07:16:52 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	8776	PASS
75	95	30	60	53.8	21768	PASS
95	95	100	100	100.0	40450	PASS
96	95	5	9	7.0	2847	PASS
173	174	0.00	2	1.5	515	PASS
174	95	50	100	84.3	34106	PASS
175	174	5	9	6.7	2280	PASS
176	174	95	101	95.8	32677	PASS
177	176	5	9	7.0	2295	PASS