

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050620\
 Data File : VN061312.D
 Acq On : 6 May 2020 16:13
 Operator : JC/MD
 Sample : PB128690TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128690TB

Quant Time: May 07 09:07:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	89527	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	156953	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	148738	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	64388	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	71723	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
35) Dibromofluoromethane	7.55	113	34050	36.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.38%	
50) Toluene-d8	10.06	98	193656	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	12.38	95	71306	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
Target Compounds						
16) Acetone	3.78	43	6438	6.367	ug/l	97
20) Methylene Chloride	4.50	84	1135	1.028	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050620\
Data File : VN061312.D
Acq On : 6 May 2020 16:13
Operator : JC/MD
Sample : PB128690TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128690TB

Quant Time: May 07 09:07:23 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

