

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
 Data File : VN067001.D
 Acq On : 6 May 2021 11:37
 Operator : JC/MD
 Sample : VN0506WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0506WBL01

Quant Time: May 07 04:05:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	453354	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	746254	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	620201	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	220187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.951	65	379897	53.24	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.48%
35) Dibromofluoromethane	7.509	113	278382	54.93	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.86%
50) Toluene-d8	10.030	98	1022007	52.96	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.92%
62) 4-Bromofluorobenzene	12.350	95	298802	45.73	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.46%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
Data File : VN067001.D
Acq On : 6 May 2021 11:37
Operator : JC/MD
Sample : VN0506WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0506WBL01

Quant Time: May 07 04:05:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

