

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
 Data File : VN067004.D
 Acq On : 6 May 2021 12:53
 Operator : JC/MD
 Sample : M2253-05 100X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41764

Quant Time: May 07 04:06:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	483724	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.520	114	812712	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	685682	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	270688	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	406550	53.40	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.80%	
35) Dibromofluoromethane	7.514	113	292243	52.95	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.90%	
50) Toluene-d8	10.030	98	1116112	53.11	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.22%	
62) 4-Bromofluorobenzene	12.350	95	304177	42.74	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 85.48%	
Target Compounds						
16) Acetone	3.760	43	8992	3.36	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
Data File : VN067004.D
Acq On : 6 May 2021 12:53
Operator : JC/MD
Sample : M2253-05 100X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41764

Quant Time: May 07 04:06:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

