

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
 Data File : VN067007.D
 Acq On : 6 May 2021 14:04
 Operator : JC/MD
 Sample : PB136189ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136189ZHE#01

Quant Time: May 07 04:07:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	461379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	776084	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	658215	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.297	152	220264	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	390724	53.80	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 107.60%			
35) Dibromofluoromethane	7.514	113	281898	53.49	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 106.98%			
50) Toluene-d8	10.030	98	1050741	52.35	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.70%			
62) 4-Bromofluorobenzene	12.350	95	300982	44.29	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 88.58%			
Target Compounds						
16) Acetone	3.760	43	20289	7.94	ug/l	93
20) Methylene Chloride	4.476	84	22726	3.51	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
Data File : VN067007.D
Acq On : 6 May 2021 14:04
Operator : JC/MD
Sample : PB136189ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136189ZHE#01

Quant Time: May 07 04:07:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

