

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
 Data File : VN067012.D
 Acq On : 6 May 2021 16:00
 Operator : JC/MD
 Sample : M2275-18
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CRT-020

Manual Integrations
 APPROVED

MMDadoda
 5/7/2021 6:25:19 PM

Quant Time: May 07 04:08:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	450557	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.518	114	760115	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	671215	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	225076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	385933	54.42	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 108.84%			
35) Dibromofluoromethane	7.512	113	274176	53.12	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 106.24%			
50) Toluene-d8	10.030	98	1033906	52.60	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 105.20%			
62) 4-Bromofluorobenzene	12.350	95	296780m	44.59	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 89.18%			
Target Compounds						
16) Acetone	3.762	43	19578	7.85	ug/l	92
20) Methylene Chloride	4.473	84	48243	7.62	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
Data File : VN067012.D
Acq On : 6 May 2021 16:00
Operator : JC/MD
Sample : M2275-18
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CRT-020

Manual Integrations
APPROVED

MMDadoda
5/7/2021 6:25:19 PM

Quant Time: May 07 04:08:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

