

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
 Data File : VN067014.D
 Acq On : 6 May 2021 16:48
 Operator : JC/MD
 Sample : M2266-08RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CRT-016RE

Manual Integrations
 APPROVED

MMDadoda
 5/7/2021 6:25:21 PM

Quant Time: May 07 04:09:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	438058	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	740150	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	643708	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	226812	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	387951	56.26	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 112.52%			
35) Dibromofluoromethane	7.512	113	271158	53.95	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 107.90%			
50) Toluene-d8	10.030	98	1031010	53.87	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.74%			
62) 4-Bromofluorobenzene	12.350	95	308538m	47.61	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.22%			
Target Compounds						Qvalue
16) Acetone	3.754	43	21274	8.77	ug/l	100
20) Methylene Chloride	4.470	84	42902	6.97	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
Data File : VN067014.D
Acq On : 6 May 2021 16:48
Operator : JC/MD
Sample : M2266-08RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CRT-016RE

Manual Integrations
APPROVED

MMDadoda
5/7/2021 6:25:21 PM

Quant Time: May 07 04:09:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

