

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
 Data File : VN067016.D
 Acq On : 6 May 2021 17:36
 Operator : JC/MD
 Sample : M2322-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-20210505

Quant Time: May 07 04:09:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	445833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	755014	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	637309	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	252560	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	378573	53.95	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 107.90%			
35) Dibromofluoromethane	7.512	113	273177	53.28	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 106.56%			
50) Toluene-d8	10.030	98	1009556	51.71	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.42%			
62) 4-Bromofluorobenzene	12.353	95	299212	45.26	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 90.52%			
Target Compounds					Qvalue	
16) Acetone	3.757	43	86282	34.95	ug/l	95
20) Methylene Chloride	4.473	84	341768	54.58	ug/l	93
52) Toluene	10.094	92	58721	4.05	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050621\
Data File : VN067016.D
Acq On : 6 May 2021 17:36
Operator : JC/MD
Sample : M2322-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-20210505

Quant Time: May 07 04:09:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

