

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
 Data File : VN072348.D
 Acq On : 06 May 2022 12:20
 Operator : JC\MD
 Sample : VN0506MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0506MBL01

Quant Time: May 09 03:11:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	206047	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	392369	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	418632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	161017	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	161270	50.982	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.960%
35) Dibromofluoromethane	8.016	113	123597	49.215	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.420%
50) Toluene-d8	10.439	98	490177	49.672	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.340%
62) 4-Bromofluorobenzene	12.727	95	190271	58.090	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	116.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
Data File : VN072348.D
Acq On : 06 May 2022 12:20
Operator : JC\MD
Sample : VN0506MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0506MBL01

Quant Time: May 09 03:11:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

