

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
 Data File : VN072358.D
 Acq On : 06 May 2022 16:35
 Operator : JC\MD
 Sample : N2725-09 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRE-1593

Quant Time: May 09 03:14:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	242575	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	459345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	502573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	209952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	184680	49.591	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.180%
35) Dibromofluoromethane	8.016	113	143179	48.699	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.400%
50) Toluene-d8	10.439	98	586381	50.757	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.520%
62) 4-Bromofluorobenzene	12.727	95	229401	59.825	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	119.640%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	4419	2.752	ug/l	# 83
25) 2-Butanone	7.334	43	2639	1.032	ug/l	92
52) Toluene	10.498	92	3521	0.444	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
Data File : VN072358.D
Acq On : 06 May 2022 16:35
Operator : JC\MD
Sample : N2725-09 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRE-1593

Quant Time: May 09 03:14:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

