

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
 Data File : VN072359.D
 Acq On : 06 May 2022 16:59
 Operator : JC\MD
 Sample : N2726-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FLOOR-DRAINING-TOTE

Quant Time: May 09 03:14:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	248284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	468366	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	517457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	241161	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	187205	49.113	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	98.220%	
35) Dibromofluoromethane	8.022	113	147229	49.112	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	98.220%	
50) Toluene-d8	10.439	98	608307	51.641	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	103.280%	
62) 4-Bromofluorobenzene	12.727	95	251896	64.426	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	128.860%	
Target Compounds						
					Qvalue	
16) Acetone	4.281	43	601903	366.258	ug/l	96
18) Methyl Acetate	4.857	43	13170	2.829	ug/l	95
20) Methylene Chloride	5.098	84	6825	2.167	ug/l #	87
24) 1,1-Dichloroethane	6.375	63	3224	0.568	ug/l #	82
25) 2-Butanone	7.333	43	31696	12.111	ug/l #	85
30) Chloroform	7.822	83	3493	0.613	ug/l #	79
37) Ethyl Acetate	7.410	43	41405	7.145	ug/l	97
40) Benzene	8.457	78	5077	0.375	ug/l	94
51) 4-Methyl-2-Pentanone	10.327	43	32907	6.019	ug/l	87
52) Toluene	10.504	92	16145	1.995	ug/l	98
67) Ethyl Benzene	11.839	91	9781	0.536	ug/l	90
68) m/p-Xylenes	11.945	106	12526	1.850	ug/l	89
69) o-Xylene	12.274	106	8925	1.360	ug/l	89
80) 1,3,5-Trimethylbenzene	13.057	105	8826	0.498	ug/l	92
84) 1,2,4-Trimethylbenzene	13.362	105	39970	2.329	ug/l	78
95) Naphthalene	15.504	128	38843	7.142	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

