

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
 Data File : VN072360.D
 Acq On : 06 May 2022 17:24
 Operator : JC\MD
 Sample : N2712-13 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GROUND-WATER

Quant Time: May 09 03:14:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	258046	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	495335	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	544669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	228064	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	197585	49.875	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.760%
35) Dibromofluoromethane	8.022	113	154930	48.867	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.740%
50) Toluene-d8	10.439	98	630732	50.629	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.260%
62) 4-Bromofluorobenzene	12.727	95	257001	62.153	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	124.300%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	3880	2.272	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
Data File : VN072360.D
Acq On : 06 May 2022 17:24
Operator : JC\MD
Sample : N2712-13 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GROUND-WATER

Quant Time: May 09 03:14:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

