

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
 Data File : VN072366.D
 Acq On : 06 May 2022 19:49
 Operator : JC\MD
 Sample : N2672-05ME
 Misc : 5.52g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 05ME

Quant Time: May 09 03:16:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.075	168	281960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	542809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	624450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	378451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	203608	47.037	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.080%
35) Dibromofluoromethane	8.010	113	156239	44.970	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.940%
50) Toluene-d8	10.439	98	712358	52.180	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.360%
62) 4-Bromofluorobenzene	12.727	95	318511	70.291	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	140.580%#
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.363	105	92355	3.429	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050622\
Data File : VN072366.D
Acq On : 06 May 2022 19:49
Operator : JC\MD
Sample : N2672-05ME
Misc : 5.52g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
05ME

Quant Time: May 09 03:16:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

