


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN050622\  
Data File : VN072346.D  
Acq On : 06 May 2022 10:26  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042622W.M  
Title : SW846 8260  
Last Update : Tue Apr 26 18:50:11 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 20.0         | 7388       | PASS                |
| 75             | 95              | 30              | 60              | 58.6         | 21627      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 36915      | PASS                |
| 96             | 95              | 5               | 9               | 5.9          | 2163       | PASS                |
| 173            | 174             | 0.00            | 2               | 0.9          | 260        | PASS                |
| 174            | 95              | 50              | 100             | 79.2         | 29253      | PASS                |
| 175            | 174             | 5               | 9               | 7.4          | 2163       | PASS                |
| 176            | 174             | 95              | 101             | 96.6         | 28259      | PASS                |
| 177            | 176             | 5               | 9               | 5.8          | 1632       | PASS                |