

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
 Data File : VN082007.D
 Acq On : 06 May 2024 14:10
 Operator : JC\MD
 Sample : P2382-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(MAY)

Quant Time: May 07 00:28:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	32305	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	194201	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	170012	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	83820	30.990	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.300%
60) 4-Bromofluorobenzene	12.847	95	78171	27.628	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.100%
63) Toluene-d8	10.571	98	244812	30.321	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.067%
Target Compounds						
15) Acetone	4.424	58	238273	943.708	ug/l	90
25) Chloroform	7.965	83	3216	0.824	ug/l #	69
30) 2-Butanone	7.483	43	101265	67.948	ug/l	95
58) 4-Methyl-2-Pentanone	10.447	43	7265	2.201	ug/l #	77
62) Toluene	10.629	91	135594	14.175	ug/l	97
67) m/p-Xylenes	12.070	106	3091	0.795	ug/l	91
68) o-Xylene	12.388	106	1393	0.362	ug/l	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
Data File : VN082007.D
Acq On : 06 May 2024 14:10
Operator : JC\MD
Sample : P2382-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(MAY)

Quant Time: May 07 00:28:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

