

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
Data File : VN082008.D
Acq On : 06 May 2024 14:48
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 07 00:28:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	31083	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	179060	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	152208	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	80394	30.892	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.967%
60) 4-Bromofluorobenzene	12.853	95	69970	27.623	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.067%
63) Toluene-d8	10.571	98	219500	30.366	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.233%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
Data File : VN082008.D
Acq On : 06 May 2024 14:48
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 07 00:28:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

