

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
 Data File : VN082009.D
 Acq On : 06 May 2024 15:34
 Operator : JC\MD
 Sample : P2401-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240503004-03-TRIP-BLANK

Quant Time: May 07 00:29:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	32314	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	191129	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	163111	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	81608	30.164	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.533%
60) 4-Bromofluorobenzene	12.847	95	74788	27.551	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.833%
63) Toluene-d8	10.571	98	236863	30.578	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.933%
Target Compounds					Qvalue	
25) Chloroform	7.965	83	18431	4.723	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
Data File : VN082009.D
Acq On : 06 May 2024 15:34
Operator : JC\MD
Sample : P2401-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240503004-03-TRIP-BLANK

Quant Time: May 07 00:29:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

