

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
 Data File : VN082000.D
 Acq On : 06 May 2024 09:35
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: May 07 00:24:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	84	0.00
2 M	Dichlorodifluoromethane	1.703	1.576	7.5	79	0.00
3 M	Chloromethane	1.975	1.847	6.5	81	0.00
4 M	Vinyl Chloride	1.866	1.786	4.3	83	0.00
5 M	Bromomethane	0.974	0.893	8.3	77	0.00
6 M	Chloroethane	1.056	1.105	-4.6	87	0.00
7 M	Trichlorofluoromethane	2.730	2.630	3.7	81	0.00
8 T	Diethyl Ether	1.322	1.244	5.9	78	0.00
9	1,1,2-Trichlorotrifluoroeth	1.872	1.706	8.9	76	0.02
10 M	1,1-Dichloroethene	1.831	1.760	3.9	83	0.00
11	Methyl Iodide	1.673	1.472	12.0	85	0.00
12	Methyl Acetate	2.293	2.456	-7.1	92	0.00
13 M	Acrolein	0.487	0.409	16.0	74	0.00
14 M	Acrylonitrile	1.046	1.083	-3.5	87	0.00
15 M	Acetone	0.234	0.261	-11.5	91	0.00
16 M	Carbon Disulfide	5.073	4.634	8.7	80	0.00
17	Allyl chloride	3.215	3.202	0.4	81	0.00
18 M	Methylene Chloride	2.083	2.066	0.8	87	-0.01
19 M	trans-1,2-Dichloroethene	1.906	1.824	4.3	85	-0.01
20 T	Diisopropyl ether	7.112	6.891	3.1	82	0.00
21 M	1,1-Dichloroethane	3.766	3.785	-0.5	84	0.00
22 M	cis-1,2-Dichloroethene	2.322	2.339	-0.7	87	0.00
23 M	tert-Butyl Alcohol	0.431	0.418	3.0	78	0.02
24 M	Methyl tert-Butyl Ether	6.601	6.314	4.3	79	0.00
25 M	Chloroform	3.623	3.608	0.4	84	0.00
26	Cyclohexane	3.257	3.185	2.2	85	0.00
27 s	1,2-Dichloroethane-d4	2.512	2.620	-4.3	89	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
29	1,1-Dichloropropene	0.447	0.402	10.1	79	0.00
30 M	2-Butanone	0.230	0.228	0.9	85	0.00
31	2,2-Dichloropropane	0.458	0.520	-13.5	102	0.00
32 M	1,1,1-Trichloroethane	0.514	0.496	3.5	84	0.00
33 M	Carbon Tetrachloride	0.417	0.383	8.2	80	0.00
34 M	Benzene	1.425	1.382	3.0	84	0.00
35	Methacrylonitrile	0.291	0.289	0.7	88	0.00
36 M	1,2-Dichloroethane	0.478	0.442	7.5	82	0.00
37 M	Trichloroethene	0.349	0.313	10.3	79	0.00
38	Methylcyclohexane	0.540	0.503	6.9	81	0.00
39 M	1,2-Dichloropropane	0.366	0.360	1.6	89	0.00
40	Dibromomethane	0.235	0.229	2.6	87	0.00
41 M	Bromodichloromethane	0.502	0.474	5.6	85	0.00
42 M	Vinyl Acetate	0.980	0.925	5.6	84	0.00
43	Ethyl Acetate	0.502	0.496	1.2	88	0.00
44	Isopropyl Acetate	0.891	0.855	4.0	84	0.00
45 T	1,4-Dioxane	0.007	0.007	0.0	84	0.00
46	Methyl methacrylate	0.407	0.398	2.2	86	0.00
47	n-amyl Acetate	0.804	0.731	9.1	81	0.00
48 M	t-1,3-Dichloropropene	0.544	0.507	6.8	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
 Data File : VN082000.D
 Acq On : 06 May 2024 09:35
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: May 07 00:24:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.576	0.557	3.3	85	0.00
50 M	1,1,2-Trichloroethane	0.319	0.322	-0.9	88	0.00
51	Ethyl methacrylate	0.601	0.539	10.3	79	0.00
52	1,3-Dichloropropane	0.585	0.558	4.6	84	0.00
53 M	Dibromochloromethane	0.347	0.311	10.4	82	0.00
54 M	1,2-Dibromoethane	0.318	0.299	6.0	82	0.00
55 M	2-Chloroethyl vinyl ether	0.301	0.314	-4.3	91	0.00
56 M	Bromoform	0.211	0.187	11.4	78	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
58 M	4-Methyl-2-Pentanone	0.583	0.576	1.2	84	0.00
59 M	2-Hexanone	0.432	0.433	-0.2	86	0.00
60 S	4-Bromofluorobenzene	0.499	0.480	3.8	85	0.00
61 M	Tetrachloroethene	0.333	0.289	13.2	73	0.00
62 M	Toluene	1.688	1.623	3.9	83	0.00
63 S	Toluene-d8	1.425	1.456	-2.2	88	0.00
64 M	Chlorobenzene	1.032	0.978	5.2	81	0.00
65	1,1,1,2-Tetrachloroethane	0.338	0.317	6.2	80	0.00
66 M	Ethyl Benzene	1.850	1.793	3.1	81	0.00
67 M	m/p-Xylenes	0.686	0.660	3.8	82	0.00
68 M	o-Xylene	0.679	0.665	2.1	82	0.00
69 M	Styrene	1.139	1.085	4.7	83	0.00
70	Isopropylbenzene	1.724	1.633	5.3	80	0.00
71 M	1,1,2,2-Tetrachloroethane	0.496	0.502	-1.2	84	0.00
72	1,2,3-Trichloropropane	0.464	0.453	2.4	85	0.00
73	Bromobenzene	0.369	0.339	8.1	81	0.00
74	n-propylbenzene	2.034	2.010	1.2	84	0.00
75	2-Chlorotoluene	1.258	1.152	8.4	79	0.00
76	1,3,5-Trimethylbenzene	1.395	1.323	5.2	80	0.00
77	t-1,4-Dichloro-2-butene	0.206	0.214	-3.9	91	0.00
78	4-Chlorotoluene	1.233	1.174	4.8	83	0.00
79	tert-butylbenzene	1.201	1.130	5.9	79	0.00
80	1,2,4-Trimethylbenzene	1.401	1.319	5.9	81	0.00
81	sec-Butylbenzene	1.684	1.616	4.0	83	0.00
82	p-Isopropyltoluene	1.332	1.326	0.5	84	0.00
83 M	1,3-Dichlorobenzene	0.664	0.651	2.0	83	0.00
84 M	1,4-Dichlorobenzene	0.668	0.650	2.7	84	0.00
85	n-Butylbenzene	1.216	1.229	-1.1	89	0.00
86 T	Hexachloroethane	0.266	0.260	2.3	84	0.00
87 M	1,2-Dichlorobenzene	0.662	0.622	6.0	81	0.00
88	1,2-Dibromo-3-Chloropropane	0.111	0.106	4.5	78	0.00
89	1,2,4-Trichlorobenzene	0.364	0.332	8.8	80	0.00
90	Hexachlorobutadiene	0.135	0.109	19.3	74	0.00
91 M	Naphthalene	1.385	1.263	8.8	78	0.00
92	1,2,3-Trichlorobenzene	0.364	0.332	8.8	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0