

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
 Data File : VN082013.D
 Acq On : 06 May 2024 17:20
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: May 07 01:12:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	72	0.00
2 M	Dichlorodifluoromethane	1.703	1.532	10.0	66	0.00
3 M	Chloromethane	1.975	2.025	-2.5	76	0.00
4 M	Vinyl Chloride	1.866	1.964	-5.3	79	0.00
5 M	Bromomethane	0.974	0.884	9.2	66	0.00
6 M	Chloroethane	1.056	1.546	-46.4#	105	0.00
7 M	Trichlorofluoromethane	2.730	2.646	3.1	70	0.00
8 T	Diethyl Ether	1.322	1.356	-2.6	74	0.00
9	1,1,2-Trichlorotrifluoroeth	1.872	1.856	0.9	71	0.00
10 M	1,1-Dichloroethene	1.831	1.737	5.1	70	0.00
11	Methyl Iodide	1.673	1.267	24.3	63	-0.01
12	Methyl Acetate	2.293	2.500	-9.0	81	0.00
13 M	Acrolein	0.487	0.188	61.4#	29#	0.00
14 M	Acrylonitrile	1.046	1.092	-4.4	76	0.00
15 M	Acetone	0.234	0.240	-2.6	72	0.00
16 M	Carbon Disulfide	5.073	4.565	10.0	68	0.00
17	Allyl chloride	3.215	3.138	2.4	69	0.01
18 M	Methylene Chloride	2.083	2.185	-4.9	79	-0.01
19 M	trans-1,2-Dichloroethene	1.906	1.826	4.2	74	0.00
20 T	Diisopropyl ether	7.112	7.473	-5.1	77	0.00
21 M	1,1-Dichloroethane	3.766	4.080	-8.3	78	0.00
22 M	cis-1,2-Dichloroethene	2.322	2.400	-3.4	77	0.00
23 M	tert-Butyl Alcohol	0.431	0.388	10.0	62	0.02
24 M	Methyl tert-Butyl Ether	6.601	6.511	1.4	70	0.00
25 M	Chloroform	3.623	3.804	-5.0	76	0.00
26	Cyclohexane	3.257	3.268	-0.3	75	0.00
27 s	1,2-Dichloroethane-d4	2.512	2.632	-4.8	77	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
29	1,1-Dichloropropene	0.447	0.444	0.7	77	0.00
30 M	2-Butanone	0.230	0.220	4.3	72	0.00
31	2,2-Dichloropropane	0.458	0.513	-12.0	88	0.00
32 M	1,1,1-Trichloroethane	0.514	0.475	7.6	70	0.00
33 M	Carbon Tetrachloride	0.417	0.391	6.2	71	0.00
34 M	Benzene	1.425	1.417	0.6	75	0.00
35	Methacrylonitrile	0.291	0.281	3.4	75	0.00
36 M	1,2-Dichloroethane	0.478	0.465	2.7	75	0.00
37 M	Trichloroethene	0.349	0.293	16.0	65	0.00
38	Methylcyclohexane	0.540	0.523	3.1	74	0.00
39 M	1,2-Dichloropropane	0.366	0.388	-6.0	84	0.00
40	Dibromomethane	0.235	0.224	4.7	75	0.00
41 M	Bromodichloromethane	0.502	0.498	0.8	78	0.00
42 M	Vinyl Acetate	0.980	0.955	2.6	76	0.00
43	Ethyl Acetate	0.502	0.481	4.2	75	0.00
44	Isopropyl Acetate	0.891	0.852	4.4	73	0.00
45 T	1,4-Dioxane	0.007	0.006	14.3	68	0.00
46	Methyl methacrylate	0.407	0.394	3.2	75	0.00
47	n-amyl Acetate	0.804	0.766	4.7	75	0.00
48 M	t-1,3-Dichloropropene	0.544	0.507	6.8	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
 Data File : VN082013.D
 Acq On : 06 May 2024 17:20
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: May 07 01:12:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.576	0.581	-0.9	78	0.00
50 M	1,1,2-Trichloroethane	0.319	0.325	-1.9	78	0.00
51	Ethyl methacrylate	0.601	0.581	3.3	75	0.00
52	1,3-Dichloropropane	0.585	0.574	1.9	76	0.00
53 M	Dibromochloromethane	0.347	0.317	8.6	73	0.00
54 M	1,2-Dibromoethane	0.318	0.312	1.9	75	0.00
55 M	2-Chloroethyl vinyl ether	0.301	0.289	4.0	73	0.00
56 M	Bromoform	0.211	0.177	16.1	65	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
58 M	4-Methyl-2-Pentanone	0.583	0.556	4.6	73	0.00
59 M	2-Hexanone	0.432	0.411	4.9	73	0.00
60 S	4-Bromofluorobenzene	0.499	0.478	4.2	76	0.00
61 M	Tetrachloroethene	0.333	0.259	22.2	58	0.00
62 M	Toluene	1.688	1.645	2.5	75	0.00
63 S	Toluene-d8	1.425	1.452	-1.9	79	0.00
64 M	Chlorobenzene	1.032	0.998	3.3	74	0.00
65	1,1,1,2-Tetrachloroethane	0.338	0.314	7.1	71	0.00
66 M	Ethyl Benzene	1.850	1.825	1.4	74	0.00
67 M	m/p-Xylenes	0.686	0.659	3.9	74	0.00
68 M	o-Xylene	0.679	0.631	7.1	70	0.00
69 M	Styrene	1.139	1.089	4.4	74	0.00
70	Isopropylbenzene	1.724	1.632	5.3	72	0.00
71 M	1,1,2,2-Tetrachloroethane	0.496	0.534	-7.7	81	0.00
72	1,2,3-Trichloropropane	0.464	0.442	4.7	75	0.00
73	Bromobenzene	0.369	0.345	6.5	74	0.00
74	n-propylbenzene	2.034	2.004	1.5	75	0.00
75	2-Chlorotoluene	1.258	1.223	2.8	75	0.00
76	1,3,5-Trimethylbenzene	1.395	1.333	4.4	72	0.00
77	t-1,4-Dichloro-2-butene	0.206	0.196	4.9	75	0.00
78	4-Chlorotoluene	1.233	1.160	5.9	74	0.00
79	tert-butylbenzene	1.201	1.149	4.3	72	0.00
80	1,2,4-Trimethylbenzene	1.401	1.330	5.1	74	0.00
81	sec-Butylbenzene	1.684	1.654	1.8	76	0.00
82	p-Isopropyltoluene	1.332	1.299	2.5	74	0.00
83 M	1,3-Dichlorobenzene	0.664	0.618	6.9	71	0.00
84 M	1,4-Dichlorobenzene	0.668	0.605	9.4	70	0.00
85	n-Butylbenzene	1.216	1.223	-0.6	79	0.00
86 T	Hexachloroethane	0.266	0.226	15.0	65	0.00
87 M	1,2-Dichlorobenzene	0.662	0.629	5.0	73	0.00
88	1,2-Dibromo-3-Chloropropane	0.111	0.092	17.1	61	0.00
89	1,2,4-Trichlorobenzene	0.364	0.309	15.1	67	0.00
90	Hexachlorobutadiene	0.135	0.102	24.4	62	0.00
91 M	Naphthalene	1.385	1.240	10.5	69	0.00
92	1,2,3-Trichlorobenzene	0.364	0.309	15.1	67	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0