



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050624\
Data File : VN081999.D
Acq On : 06 May 2024 08:34
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Sat Apr 27 04:37:27 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.0	9494	PASS
75	95	30	60	59.3	22501	PASS
95	95	100	100	100.0	37941	PASS
96	95	5	9	5.9	2251	PASS
173	174	0.00	2	0.9	230	PASS
174	95	50	100	65.4	24795	PASS
175	174	5	9	6.4	1575	PASS
176	174	95	101	96.8	24011	PASS
177	176	5	9	4.2	1012	FAIL*