

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050625\
 Data File : VN086510.D
 Acq On : 06 May 2025 13:22
 Operator : JC\MD
 Sample : Q1947-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-O

Quant Time: May 07 02:55:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	323480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	313188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	139925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121653	48.313	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.620%
35) Dibromofluoromethane	8.165	113	84532	56.305	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.620%
50) Toluene-d8	10.565	98	410428	51.152	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.300%
62) 4-Bromofluorobenzene	12.847	95	151084	51.624	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.240%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	30630	30.165	ug/l	100
20) Methylene Chloride	5.277	84	7564	3.249	ug/l	95
43) Isopropyl Acetate	8.683	43	25141	2.624	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050625\
Data File : VN086510.D
Acq On : 06 May 2025 13:22
Operator : JC\MD
Sample : Q1947-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-O

Quant Time: May 07 02:55:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

