

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050719\
 Data File : VN055437.D
 Acq On : 7 May 2019 15:04
 Operator : JC/SP
 Sample : VN0507MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0507MBS01

Quant Time: May 07 15:26:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	341995	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	568146	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	478946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	182622	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	243748	54.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.96%	
35) Dibromofluoromethane	7.59	113	186950	50.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.26%	
50) Toluene-d8	10.09	98	660566	49.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.12%	
62) 4-Bromofluorobenzene	12.40	95	220872	52.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.96%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.85	85	76116	18.165	ug/l	98
3) Chloromethane	2.06	50	88223	17.004	ug/l	97
4) Vinyl Chloride	2.19	62	77183	18.454	ug/l	100
5) Bromomethane	2.57	94	43571	19.668	ug/l	95
6) Chloroethane	2.71	64	42913	18.940	ug/l	98
7) Trichlorofluoromethane	3.02	101	115835	19.477	ug/l	100
8) Diethyl Ether	3.42	74	44242	21.263	ug/l	92
9) 1,1,2-Trichlorotrifluoroet	3.76	101	69304	19.569	ug/l	99
10) Methyl Iodide	3.96	142	91272	18.313	ug/l	99
11) Tert butyl alcohol	4.80	59	34557	134.908	ug/l	# 73
12) 1,1-Dichloroethene	3.74	96	66785	19.769	ug/l	99
13) Acrolein	3.61	56	11607	58.022	ug/l	98
14) Allyl chloride	4.33	41	142206	20.324	ug/l	99
15) Acrylonitrile	5.00	53	174077	120.903	ug/l	98
16) Acetone	3.82	43	142476	102.202	ug/l	100
17) Carbon Disulfide	4.06	76	142487	16.161	ug/l	97
18) Methyl Acetate	4.33	43	89716	25.920	ug/l	100
19) Methyl tert-butyl Ether	5.06	73	222106	23.386	ug/l	99
20) Methylene Chloride	4.56	84	81373	19.124	ug/l	96
21) trans-1,2-Dichloroethene	5.05	96	70426	19.684	ug/l	95
22) Diisopropyl ether	5.96	45	313231	22.531	ug/l	94
23) Vinyl Acetate	5.90	43	1040433	114.190	ug/l	99
24) 1,1-Dichloroethane	5.85	63	155439	20.985	ug/l	100
25) 2-Butanone	6.84	43	215396	120.817	ug/l	99
26) 2,2-Dichloropropane	6.83	77	102802	20.098	ug/l	99
27) cis-1,2-Dichloroethene	6.83	96	85783	21.083	ug/l	99
28) Bromochloromethane	7.20	49	79963	21.416	ug/l	97
29) Tetrahydrofuran	7.22	42	144563	126.633	ug/l	99
30) Chloroform	7.37	83	146997	20.935	ug/l	97
31) Cyclohexane	7.66	56	131622	18.703	ug/l	95
32) 1,1,1-Trichloroethane	7.57	97	117667	20.648	ug/l	99
36) 1,1-Dichloropropene	7.79	75	99424	18.270	ug/l	99
37) Ethyl Acetate	6.94	43	91307	21.892	ug/l	# 79
38) Carbon Tetrachloride	7.78	117	96208	18.327	ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050719\
 Data File : VN055437.D
 Acq On : 7 May 2019 15:04
 Operator : JC/SP
 Sample : VN0507MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0507MBS01

Quant Time: May 07 15:26:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.08	83	110742	19.140	ug/l	95
40) Benzene	8.04	78	318028	19.824	ug/l	99
41) Methacrylonitrile	7.18	41	51143	23.061	ug/l #	86
42) 1,2-Dichloroethane	8.13	62	114436	20.140	ug/l	100
43) Isopropyl Acetate	8.17	43	160847	23.997	ug/l #	87
44) Trichloroethene	8.84	130	76278	18.651	ug/l	97
45) 1,2-Dichloropropane	9.12	63	95240	20.584	ug/l	100
46) Dibromomethane	9.21	93	51617	20.409	ug/l	99
47) Bromodichloromethane	9.40	83	110333	20.335	ug/l	99
48) Methyl methacrylate	9.20	41	72579	22.026	ug/l	98
49) 1,4-Dioxane	9.20	88	16301	478.564	ug/l	96
51) 4-Methyl-2-Pentanone	9.99	43	461310	122.242	ug/l	97
52) Toluene	10.16	92	186835	20.298	ug/l	98
53) t-1,3-Dichloropropene	10.38	75	96034	20.389	ug/l	98
54) cis-1,3-Dichloropropene	9.84	75	118838	20.236	ug/l	96
55) 1,1,2-Trichloroethane	10.56	97	74737	21.865	ug/l	99
56) Ethyl methacrylate	10.43	69	96860	23.023	ug/l	99
57) 1,3-Dichloropropane	10.71	76	127011	21.182	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.70	63	100199	85.338	ug/l	99
59) 2-Hexanone	10.75	43	281577	119.253	ug/l	99
60) Dibromochloromethane	10.90	129	74803	21.412	ug/l	100
61) 1,2-Dibromoethane	11.01	107	68764	21.765	ug/l	98
64) Tetrachloroethene	10.63	164	77703	19.839	ug/l	98
65) Chlorobenzene	11.43	112	188793	20.161	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.51	131	74216	20.241	ug/l	99
67) Ethyl Benzene	11.51	91	341445	20.192	ug/l	98
68) m/p-Xylenes	11.62	106	246263	40.801	ug/l	99
69) o-Xylene	11.95	106	122870	20.538	ug/l	99
70) Styrene	11.96	104	199232	21.690	ug/l	100
71) Bromoform	12.13	173	45048	22.141	ug/l #	94
73) Isopropylbenzene	12.25	105	333466	21.122	ug/l	99
74) N-amyl acetate	12.07	43	99551	21.241	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.50	83	89828	23.116	ug/l	100
76) 1,2,3-Trichloropropane	12.55	75	108166	30.904	ug/l #	100
77) Bromobenzene	12.53	156	77907	21.087	ug/l	97
78) n-propylbenzene	12.59	91	362412	19.319	ug/l	99
79) 2-Chlorotoluene	12.67	91	222822	20.967	ug/l	100
80) 1,3,5-Trimethylbenzene	12.73	105	277973	19.416	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.30	75	19593	20.332	ug/l	87
82) 4-Chlorotoluene	12.77	91	211032	19.348	ug/l	99
83) tert-Butylbenzene	12.99	119	238787	21.929	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	263686	20.593	ug/l	98
85) sec-Butylbenzene	13.17	105	312273	20.067	ug/l	100
86) p-Isopropyltoluene	13.29	119	264781	20.702	ug/l	100
87) 1,3-Dichlorobenzene	13.28	146	123330	20.182	ug/l	97
88) 1,4-Dichlorobenzene	13.36	146	118323	20.560	ug/l	97
89) n-Butylbenzene	13.61	91	184989	20.354	ug/l	97
90) Hexachloroethane	13.87	117	46793	22.231	ug/l	97
91) 1,2-Dichlorobenzene	13.65	146	124988	20.636	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	14.27	75	12126	25.239	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050719\
 Data File : VN055437.D
 Acq On : 7 May 2019 15:04
 Operator : JC/SP
 Sample : VN0507MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0507MBS01

Quant Time: May 07 15:26:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.91	180	38525	18.377	ug/l	98
94) Hexachlorobutadiene	15.01	225	45887	23.706	ug/l	96
95) Naphthalene	15.13	128	84498	18.866	ug/l	98
96) 1,2,3-Trichlorobenzene	15.31	180	42888	19.898	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050719\
 Data File : VN055437.D
 Acq On : 7 May 2019 15:04
 Operator : JC/SP
 Sample : VN0507MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 Client Sample ID :
 VN0507MBS01

Quant Time: May 07 15:26:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration









#1
 Pentafluorobenzene
 Concen: 50.000 ug/l
 RT: 7.67 min Scan# 1834
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampled : VN0507MBS01


Tgt Ion: 168 Resp: 341995
 Ion Ratio Lower Upper
 168 100
 99 54.6 43.6 65.4



#2
 Dichlorodifluoromethane
 Concen: 18.165 ug/l
 RT: 1.85 min Scan# 25
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 85 Resp: 76116
 Ion Ratio Lower Upper
 85 100
 87 31.2 16.1 48.3









#3
 Chloromethane
 Concen: 17.004 ug/l
 RT: 2.06 min Scan# 89
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampleId : VN0507MBS01


Tgt Ion	Resp	Lower	Upper
50	100		
52	33.5	25.7	38.5



#4
 Vinyl Chloride
 Concen: 18.454 ug/l
 RT: 2.19 min Scan# 129
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
62	100		
64	31.8	25.5	38.3









#5
 Bromomethane
 Concen: 19.668 ug/l
 RT: 2.57 min Scan# 248
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion: 94 Resp: 43571
 Ion Ratio Lower Upper
 94 100
 96 89.8 76.0 114.0



#6
 Chloroethane
 Concen: 18.940 ug/l
 RT: 2.71 min Scan# 291
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 64 Resp: 42913
 Ion Ratio Lower Upper
 64 100
 66 32.0 25.0 37.4









#7
 Trichlorofluoromethane
 Concen: 19.477 ug/l
 RT: 3.02 min Scan# 389
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion:101 Resp: 115835
 Ion Ratio Lower Upper
 101 100
 103 66.0 52.6 79.0



#8
 Diethyl Ether
 Concen: 21.263 ug/l
 RT: 3.42 min Scan# 511
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 74 Resp: 44242
 Ion Ratio Lower Upper
 74 100
 45 126.0 58.9 176.6









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 19.569 ug/l
 RT: 3.76 min Scan# 619
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
101	69304		
101	100		
85	45.8	35.7	53.5
151	78.1	63.3	94.9



#10
 Methyl Iodide
 Concen: 18.313 ug/l
 RT: 3.96 min Scan# 680
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
142	91272		
142	100		
127	43.5	34.4	51.6
141	14.8	11.5	17.3









#11
 Tert butyl alcohol
 Concen: 134.908 ug/l
 RT: 4.80 min Scan# 940
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
59	100		
57	0.0	8.2	12.2#



#12
 1,1-Dichloroethene
 Concen: 19.769 ug/l
 RT: 3.74 min Scan# 612
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
96	100		
61	175.9	140.9	211.3
98	64.3	50.0	75.0









#13
 Acrolein
 Concen: 58.022 ug/l
 RT: 3.61 min Scan# 572
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
56	11607		
55	70.1	57.4	86.2



#14
 Allyl chloride
 Concen: 20.324 ug/l
 RT: 4.33 min Scan# 794
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
41	142206		
39	66.9	53.0	79.6
76	28.0	22.7	34.1









#15
 Acrylonitrile
 Concen: 120.903 ug/l
 RT: 5.00 min Scan# 1002
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
53	174077		
52	81.1	66.5	99.7
51	36.8	29.6	44.4



#16
 Acetone
 Concen: 102.202 ug/l
 RT: 3.82 min Scan# 638
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
43	142476		
58	29.6	23.6	35.4









#17
 Carbon Disulfide
 Concen: 16.161 ug/l
 RT: 4.06 min Scan# 711
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
76	142487		
78	8.6	7.6	11.4



#18
 Methyl Acetate
 Concen: 25.920 ug/l
 RT: 4.33 min Scan# 796
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
43	89716		
74	19.8	15.8	23.8









#19
 Methyl tert-butyl Ether
 Concen: 23.386 ug/l
 RT: 5.06 min Scan# 1021
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampled : VN0507MBS01


Tgt Ion: 73 Resp: 222106
 Ion Ratio Lower Upper
 73 100
 57 23.5 19.1 28.7



#20
 Methylene Chloride
 Concen: 19.124 ug/l
 RT: 4.56 min Scan# 866
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 84 Resp: 81373
 Ion Ratio Lower Upper
 84 100
 49 146.9 112.4 168.6
 51 43.6 34.6 51.8
 86 62.9 51.9 77.9









#21
 trans-1,2-Dichloroethene
 Concen: 19.684 ug/l
 RT: 5.05 min Scan# 1018
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01

Tgt Ion: 96 Resp: 70426




Ion	Ratio	Lower	Upper
96	100		
61	151.5	126.2	189.4
98	59.7	52.2	78.2




#22
 Diisopropyl ether
 Concen: 22.531 ug/l
 RT: 5.96 min Scan# 1301
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 45 Resp: 313231

Ion	Ratio	Lower	Upper
45	100		
43	64.1	46.6	70.0
87	21.1	18.1	27.1
59	9.8	8.2	12.2









#23
 Vinyl Acetate
 Concen: 114.190 ug/l
 RT: 5.90 min Scan# 1284
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion: 43 Resp: 1040433
 Ion Ratio Lower Upper
 43 100
 86 7.9 6.6 10.0



#24
 1,1-Dichloroethane
 Concen: 20.985 ug/l
 RT: 5.85 min Scan# 1269
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 63 Resp: 155439
 Ion Ratio Lower Upper
 63 100
 98 5.8 3.0 9.0
 100 3.5 1.8 5.4









#25
 2-Butanone
 Concen: 120.817 ug/l
 RT: 6.84 min Scan# 1577
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion: 43 Resp: 215396
 Ion Ratio Lower Upper
 43 100
 72 22.2 18.0 27.0



#26
 2,2-Dichloropropane
 Concen: 20.098 ug/l
 RT: 6.83 min Scan# 1572
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 77 Resp: 102802
 Ion Ratio Lower Upper
 77 100
 97 22.4 11.4 34.2









#27
 cis-1,2-Dichloroethene
 Concen: 21.083 ug/l
 RT: 6.83 min Scan# 1574
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
96	85783		
96	100		
61	159.4	0.0	316.2
98	62.2	0.0	126.6



#28
 Bromochloromethane
 Concen: 21.416 ug/l
 RT: 7.20 min Scan# 1687
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
49	79963		
49	100		
129	1.5	0.0	2.4
130	62.3	52.0	78.0









#29
 Tetrahydrofuran
 Concen: 126.633 ug/l
 RT: 7.22 min Scan# 1693
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01

Tgt Ion: 42 Resp: 144563




Ion	Ratio	Lower	Upper
42	100		
72	35.2	29.0	43.6
71	34.2	27.1	40.7




#30
 Chloroform
 Concen: 20.935 ug/l
 RT: 7.37 min Scan# 1742
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 83 Resp: 146997

Ion	Ratio	Lower	Upper
83	100		
85	66.5	51.4	77.2









#31
 Cyclohexane
 Concen: 18.703 ug/l
 RT: 7.66 min Scan# 1830
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
56	131622		
69	37.7	25.2	37.8
84	77.9	61.1	91.7



#32
 1,1,1-Trichloroethane
 Concen: 20.648 ug/l
 RT: 7.57 min Scan# 1804
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
97	117667		
99	64.5	51.8	77.8
61	50.3	39.0	58.6









#33
 1,2-Dichloroethane-d4
 Concen: 54.480 ug/l
 RT: 8.03 min Scan# 1946
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01




Tgt Ion	Resp	Lower	Upper
65	100		
67	53.6	0.0	106.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
114	100		
63	22.9	0.0	45.2
88	15.6	0.0	31.6









#35
 Dibromofluoromethane
 Concen: 50.132 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04


Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
113	186950		
113	100		
111	104.3	82.3	123.5
192	20.6	17.0	25.6



#36
 1,1-Dichloropropene
 Concen: 18.270 ug/l
 RT: 7.79 min Scan# 1872
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
75	99424		
75	100		
110	33.8	16.8	50.4
77	31.0	25.0	37.6









#37
 Ethyl Acetate
 Concen: 21.892 ug/l
 RT: 6.94 min Scan# 1606
 Delta R.T. 0.01 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampled : VN0507MBS01


Tgt Ion	Resp	Lower	Upper
43	100		
61	0.0	10.4	15.6#
70	9.1	6.6	10.0



#38
 Carbon Tetrachloride
 Concen: 18.327 ug/l
 RT: 7.78 min Scan# 1867
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
117	100		
119	99.0	77.0	115.6
121	31.1	23.8	35.8









#39
 Methylcyclohexane
 Concen: 19.140 ug/l
 RT: 9.08 min Scan# 2273
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampled : VN0507MBS01


Tgt Ion	Resp	Lower	Upper
83	110742		
55	90.6	67.8	101.6
98	44.3	36.2	54.4



#40
 Benzene
 Concen: 19.824 ug/l
 RT: 8.04 min Scan# 1950
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
78	318028		
77	24.4	19.0	28.6









#41
 Methacrylonitrile
 Concen: 23.061 ug/l
 RT: 7.18 min Scan# 1681
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01

Tgt Ion	Resp	Lower	Upper
41	100		
39	37.6	48.5	72.7#
67	71.9	54.9	82.3
52	33.3	24.9	37.3




Abundance Ion 41.00 (40.70 to 41.70): VNC
 Ion 39.00 (38.70 to 39.70): VNC
 Ion 67.00 (66.70 to 67.70): VNC
 Ion 52.00 (51.70 to 52.70): VNC






#42
 1,2-Dichloroethane
 Concen: 20.140 ug/l
 RT: 8.13 min Scan# 1976
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
62	100		
98	8.2	0.0	16.4



Abundance Ion 61.90 (61.60 to 62.60): VNC
 Ion 97.90 (97.60 to 98.60): VNC









#43
 Isopropyl Acetate
 Concen: 23.997 ug/l
 RT: 8.17 min Scan# 1988
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
43	160847		
61	16.9	20.8	31.2#
87	10.3	8.5	12.7



#44
 Trichloroethene
 Concen: 18.651 ug/l
 RT: 8.84 min Scan# 2197
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
130	76278		
95	104.7	0.0	203.4









#45
 1,2-Dichloropropane
 Concen: 20.584 ug/l
 RT: 9.12 min Scan# 2285
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampleID : VN0507MBS01


Tgt Ion	Resp	Lower	Upper
63	100		
65	29.7	23.9	35.9



#46
 Dibromomethane
 Concen: 20.409 ug/l
 RT: 9.21 min Scan# 2313
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
93	100		
95	81.5	65.7	98.5
174	100.0	81.5	122.3









#47
 Bromodichloromethane
 Concen: 20.335 ug/l
 RT: 9.40 min Scan# 2373
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
83	110333		
85	63.6	50.1	75.1
127	8.5	6.7	10.1



#48
 Methyl methacrylate
 Concen: 22.026 ug/l
 RT: 9.20 min Scan# 2310
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
41	72579		
69	74.0	58.6	87.8
39	57.4	43.8	65.6









#49
 1,4-Dioxane
 Concen: 478.564 ug/l
 RT: 9.20 min Scan# 2311
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
88	16301		
88	100		
43	35.9	33.3	49.9
58	76.6	62.7	94.1



#50
 Toluene-d8
 Concen: 49.056 ug/l
 RT: 10.09 min Scan# 2587
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
98	660566		
98	100		
100	64.3	51.7	77.5









#51
 4-Methyl-2-Pentanone
 Concen: 122.242 ug/l
 RT: 9.99 min Scan# 2554
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion: 43 Resp: 461310
 Ion Ratio Lower Upper
 43 100
 58 34.1 28.5 42.7



#52
 Toluene
 Concen: 20.298 ug/l
 RT: 10.16 min Scan# 2607
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 92 Resp: 186835
 Ion Ratio Lower Upper
 92 100
 91 171.8 140.1 210.1









#53
 t-1,3-Dichloropropene
 Concen: 20.389 ug/l
 RT: 10.38 min Scan# 2677
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion: 75 Resp: 96034
 Ion Ratio Lower Upper
 75 100
 77 31.6 24.4 36.6



#54
 cis-1,3-Dichloropropene
 Concen: 20.236 ug/l
 RT: 9.84 min Scan# 2509
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 75 Resp: 118838
 Ion Ratio Lower Upper
 75 100
 77 31.7 24.2 36.2
 39 59.2 44.2 66.2









#55
 1,1,2-Trichloroethane
 Concen: 21.865 ug/l
 RT: 10.56 min Scan# 2733
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
97	100		
83	90.6	72.8	109.2
85	59.1	47.0	70.6
99	62.2	49.0	73.4



#56
 Ethyl methacrylate
 Concen: 23.023 ug/l
 RT: 10.43 min Scan# 2692
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
69	100		
41	78.9	62.7	94.1
39	41.7	32.0	48.0









#57
 1,3-Dichloropropane
 Concen: 21.182 ug/l
 RT: 10.71 min Scan# 2779
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampled : VN0507MBS01


Tgt Ion	Resp	Lower	Upper
76	127011		
78	32.6	25.6	38.4



#58
 2-Chloroethyl Vinyl ether
 Concen: 85.338 ug/l
 RT: 9.70 min Scan# 2464
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
63	100199		
106	23.1	19.0	28.4









#59
 2-Hexanone
 Concen: 119.253 ug/l
 RT: 10.75 min Scan# 2793
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampled : VN0507MBS01


Tgt Ion	Resp	Lower	Upper
43	100		
58	48.9	24.9	74.6



#60
 Dibromochloromethane
 Concen: 21.412 ug/l
 RT: 10.90 min Scan# 2839
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
129	100		
127	76.0	38.1	114.5









#61
 1,2-Dibromoethane
 Concen: 21.765 ug/l
 RT: 11.01 min Scan# 2872
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampled : VN0507MBS01


Tgt Ion	Resp	Lower	Upper
107	100		
109	91.8	75.0	112.6



#62
 4-Bromofluorobenzene
 Concen: 52.483 ug/l
 RT: 12.40 min Scan# 3305
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
95	100		
174	81.6	0.0	166.6
176	78.7	0.0	157.8









#63
 Chlorobenzene-d5
 Concen: 50.000 ug/l
 RT: 11.41 min Scan# 2996
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument : MSVOA_N
 ClientSampled : VN0507MBS01


Tgt Ion	Resp	Lower	Upper
117	478946		
82	57.9	46.2	69.4
119	32.6	26.6	40.0



#64
 Tetrachloroethene
 Concen: 19.839 ug/l
 RT: 10.63 min Scan# 2755
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
164	77703		
166	124.6	102.3	153.5
129	93.1	75.0	112.6
131	88.3	72.7	109.1









#65
 Chlorobenzene
 Concen: 20.161 ug/l
 RT: 11.43 min Scan# 3004
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Instrument : MSVOA_N
 ClientSampled : VN0507MBS01

Tgt Ion: 112 Resp: 188793




Ion	Ratio	Lower	Upper
112	100		
114	31.7	25.3	37.9




#66
 1,1,1,2-Tetrachloroethane
 Concen: 20.241 ug/l
 RT: 11.51 min Scan# 3027
 Delta R.T. -0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 131 Resp: 74216

Ion	Ratio	Lower	Upper
131	100		
133	95.5	47.5	142.6
119	66.3	32.7	98.1









#67
Ethyl Benzene
Concen: 20.192 ug/l
RT: 11.51 min Scan# 3028
Delta R.T. 0.00 min
Lab File: VN055437.D
Acq: 7 May 2019 15:04




Instrument :
MSVOA_N
ClientSampled :
VN0507MBS01


Tgt Ion: 91 Resp: 341445
Ion Ratio Lower Upper
91 100
106 29.8 24.6 36.8



#68
m/p-Xylenes
Concen: 40.801 ug/l
RT: 11.62 min Scan# 3063
Delta R.T. 0.00 min
Lab File: VN055437.D
Acq: 7 May 2019 15:04

Tgt Ion: 106 Resp: 246263
Ion Ratio Lower Upper
106 100
91 209.8 169.5 254.3









#69
 o-Xylene
 Concen: 20.538 ug/l
 RT: 11.95 min Scan# 3164
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion:106 Resp: 122870
 Ion Ratio Lower Upper
 106 100
 91 225.1 111.5 334.5



#70
 Styrene
 Concen: 21.690 ug/l
 RT: 11.96 min Scan# 3169
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion:104 Resp: 199232
 Ion Ratio Lower Upper
 104 100
 78 52.8 42.2 63.2
 103 56.3 44.6 67.0









#71
 Bromoform
 Concen: 22.141 ug/l
 RT: 12.13 min Scan# 3220
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
173	45048		
175	51.2	23.6	71.0
254	0.0	0.0	0.0



#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.34 min Scan# 3598
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
152	182622		
115	57.7	29.0	87.0
150	164.5	0.0	345.6









#73
 Isopropylbenzene
 Concen: 21.122 ug/l
 RT: 12.25 min Scan# 3258
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
105	100		
120	25.5	12.9	38.7



#74
 N-nyl acetate
 Concen: 21.241 ug/l
 RT: 12.07 min Scan# 3202
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
43	100		
70	35.4	27.9	41.9
55	23.8	18.2	27.2
61	20.8	16.9	25.3









#75
 1,1,2,2-Tetrachloroethane
 Concen: 23.116 ug/l
 RT: 12.50 min Scan# 3337
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
83	100		
131	9.9	5.1	15.2
85	64.5	32.3	96.8



#76
 1,2,3-Trichloropropane
 Concen: 30.904 ug/l
 RT: 12.55 min Scan# 3353
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
75	100		
77	0.0	0.0	0.0









#77
 Bromobenzene
 Concen: 21.087 ug/l
 RT: 12.53 min Scan# 3344
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
156	77907		
77	204.9	99.6	298.8
158	98.2	49.3	147.8



#78
 n-propylbenzene
 Concen: 19.319 ug/l
 RT: 12.59 min Scan# 3364
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
91	362412		
120	22.4	11.0	33.0









#79
 2-Chlorotoluene
 Concen: 20.967 ug/l
 RT: 12.67 min Scan# 3390
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion: 91 Resp: 222822
 Ion Ratio Lower Upper
 91 100
 126 33.4 16.6 49.7



#80
 1,3,5-Trimethylbenzene
 Concen: 19.416 ug/l
 RT: 12.73 min Scan# 3408
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 105 Resp: 277973
 Ion Ratio Lower Upper
 105 100
 120 47.0 23.4 70.0









#81
 trans-1,4-Dichloro-2-butene
 Concen: 20.332 ug/l
 RT: 12.30 min Scan# 3274
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
75	19593		
75	100		
53	126.4	86.3	129.5
89	43.8	34.6	51.8



#82
 4-Chlorotoluene
 Concen: 19.348 ug/l
 RT: 12.77 min Scan# 3421
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
91 <td>211032</td> <td></td> <td></td>	211032		
91 <td>100</td> <td></td> <td></td>	100		
126	32.3	16.6	49.7









#83
 tert-Butylbenzene
 Concen: 21.929 ug/l
 RT: 12.99 min Scan# 3489
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
119	100		
91	69.2	34.3	102.9
134	24.1	13.1	39.3



#84
 1,2,4-Trimethylbenzene
 Concen: 20.593 ug/l
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
105	100		
120	43.3	22.1	66.5









#85
 sec-Butylbenzene
 Concen: 20.067 ug/l
 RT: 13.17 min Scan# 3545
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
105	312273		
134	19.4	9.6	28.8



#86
 p-Isopropyltoluene
 Concen: 20.702 ug/l
 RT: 13.29 min Scan# 3581
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
119	264781		
134	26.0	13.0	38.9
91	24.5	12.2	36.6









#87
 1,3-Dichlorobenzene
 Concen: 20.182 ug/l
 RT: 13.28 min Scan# 3579
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
146	123330		
111	38.8	20.3	60.8
148	62.3	32.5	97.5



#88
 1,4-Dichlorobenzene
 Concen: 20.560 ug/l
 RT: 13.36 min Scan# 3604
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
146	118323		
111	40.7	19.5	58.5
148	62.6	32.5	97.4









#89
 n-Butylbenzene
 Concen: 20.354 ug/l
 RT: 13.61 min Scan# 3682
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
91	184989		
92	50.2	26.1	78.3
134	23.5	12.4	37.4



#90
 Hexachloroethane
 Concen: 22.231 ug/l
 RT: 13.87 min Scan# 3763
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
117	46793		
201	81.0	41.9	125.8









#91
 1,2-Dichlorobenzene
 Concen: 20.636 ug/l
 RT: 13.65 min Scan# 3694
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0507MBS01

Tgt Ion: 146 Resp: 124988




Ion	Ratio	Lower	Upper
146	100		
111	41.0	20.3	61.0
148	64.2	32.4	97.2




#92
 1,2-Dibromo-3-Chloropropane
 Concen: 25.239 ug/l
 RT: 14.27 min Scan# 3886
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion: 75 Resp: 12126

Ion	Ratio	Lower	Upper
75	100		
155	82.3	41.5	124.6
157	104.5	54.0	162.2









#93
 1,2,4-Trichlorobenzene
 Concen: 18.377 ug/l
 RT: 14.91 min Scan# 4085
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04




Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01


Tgt Ion	Resp	Lower	Upper
180	38525		
182	95.1	48.3	144.9
145	29.5	14.2	42.6



#94
 Hexachlorobutadiene
 Concen: 23.706 ug/l
 RT: 15.01 min Scan# 4116
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
225	45887		
223	63.5	30.3	90.8
227	65.5	30.9	92.7









#95
 Naphthalene
 Concen: 18.866 ug/l
 RT: 15.13 min Scan# 4155
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Instrument :
 MSVOA_N
 ClientSampled :
 VN0507MBS01

Tgt Ion	Resp	Lower	Upper
128	84498		
127	12.0	10.7	16.1
129	10.5	8.6	13.0



#96
 1,2,3-Trichlorobenzene
 Concen: 19.898 ug/l
 RT: 15.31 min Scan# 4211
 Delta R.T. 0.00 min
 Lab File: VN055437.D
 Acq: 7 May 2019 15:04

Tgt Ion	Resp	Lower	Upper
180	42888		
182	95.4	48.3	144.9
145	32.3	15.2	45.6

