

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050719\
 Data File : VN055455.D
 Acq On : 7 May 2019 23:28
 Operator : JC/SP
 Sample : K2717-03
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P026-SS001-0612-01

Quant Time: May 08 08:08:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	430070	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	669107	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	544543	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	190004	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	280807	49.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.82%	
35) Dibromofluoromethane	7.59	113	209697	47.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.50%	
50) Toluene-d8	10.09	98	787527	49.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.32%	
62) 4-Bromofluorobenzene	12.40	95	247047	49.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.70%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050719\
Data File : VN055455.D
Acq On : 7 May 2019 23:28
Operator : JC/SP
Sample : K2717-03
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P026-SS001-0612-01

Quant Time: May 08 08:08:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN055455.D

Acq: 7 May 2019 23:28

Instrument :

MSVOA_N

ClientSampleId :


P026-SS001-0612-01

Tot Ion:168 Resp: 430070

Ion Ratio Lower Upper

168 100

99 55.4 43.6 65.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 49.910 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN055455.D


Acq: 7 May 2019 23:28

Tgt Ion: 65 Resp: 280807

Ion Ratio Lower Upper

65 100

67 52.9 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO


8.03


100000

50000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN055455.D

Acq: 7 May 2019 23:28

Instrument :

MSVOA_N

ClientSampleId :

P026-SS001-0612-01


Tot Ion:114 Resp: 669107

Ion Ratio Lower Upper

114 100

63 23.7 0.0 45.2




88 16.7 0.0 31.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 47.747 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN055455.D

Acq: 7 May 2019 23:28


Tgt Ion:113 Resp: 209697

Ion Ratio Lower Upper

113 100

111 102.7 82.3 123.5



192 20.4 17.0 25.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.660 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN055455.D

Acq: 7 May 2019 23:28

Instrument :

MSVOA_N

ClientSampleId :


P026-SS001-0612-01

Tot Ion: 98 Resp: 787527

Ion Ratio Lower Upper

98 100

100 63.8 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VN055286.D (-2570) (-)

Ion 100.00 (99.70 to 100.70): VN055286.D (-2570) (-)

10.09

400000



300000

200000

100000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 49.845 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN055455.D

Acq: 7 May 2019 23:28


Tgt Ion: 95 Resp: 247047

Ion Ratio Lower Upper

95 100

174 81.6 0.0 166.6

176 78.8 0.0 157.8



Abundance Ion 94.90 (94.60 to 95.60): VN055286.D (-3291) (-)

Ion 173.80 (173.50 to 174.50): VN055286.D (-3291) (-)

Ion 175.80 (175.50 to 176.50): VN055286.D (-3291) (-)

12.40


150000


100000

50000

0

Time-->





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN055455.D

Acq: 7 May 2019 23:28

Instrument :

MSVOA_N

ClientSampleId :

P026-SS001-0612-01


Tot Ion:117 Resp: 544543

Ion Ratio Lower Upper

117 100

82 58.3 46.2 69.4




119 32.0 26.6 40.0



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VN

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. 0.00 min

Lab File: VN055455.D

Acq: 7 May 2019 23:28


Tgt Ion:152 Resp: 190004

Ion Ratio Lower Upper

152 100

115 59.8 29.0 87.0

150 159.2 0.0 345.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

