

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050719\
 Data File : VN055461.D
 Acq On : 8 May 2019 2:00
 Operator : JC/SP
 Sample : K2717-08
 Misc : 6.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P026-SS002-0612-01

Quant Time: May 08 08:19:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	430975	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	664024	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	554921	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	187628	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	279194	49.52	ug/l	0.00
Spiked Amount			Recovery	=	99.04%	
35) Dibromofluoromethane	7.59	113	208689	47.88	ug/l	0.00
Spiked Amount			Recovery	=	95.76%	
50) Toluene-d8	10.09	98	789292	50.15	ug/l	0.00
Spiked Amount			Recovery	=	100.30%	
62) 4-Bromofluorobenzene	12.40	95	244983	49.81	ug/l	0.00
Spiked Amount			Recovery	=	99.62%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050719\
Data File : VN055461.D
Acq On : 8 May 2019 2:00
Operator : JC/SP
Sample : K2717-08
Misc : 6.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P026-SS002-0612-01

Quant Time: May 08 08:19:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN055461.D

Acq: 8 May 2019 2:00

Instrument :

MSVOA_N

ClientSampleId :


P026-SS002-0612-01

Tot Ion:168 Resp: 430975

Ion Ratio Lower Upper

168 100

99 54.1 43.6 65.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 49.519 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN055461.D


Acq: 8 May 2019 2:00

Tgt Ion: 65 Resp: 279194

Ion Ratio Lower Upper

65 100

67 53.6 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO


8.03


100000

50000

0

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN055461.D
 Acq: 8 May 2019 2:00

Instrument :
 MSVOA_N
 ClientSampleId :
 P026-SS002-0612-01

Tot Ion	Ratio	Lower	Upper
114	100		
63	23.2	0.0	45.2
88	16.5	0.0	31.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 47.882 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN055461.D
 Acq: 8 May 2019 2:00

Tgt Ion	Ratio	Lower	Upper
113	100		
111	102.8	82.3	123.5
192	20.6	17.0	25.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.153 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN055461.D

Acq: 8 May 2019 2:00

Instrument :

MSVOA_N

ClientSampleId :





P026-SS002-0612-01

Tot Ion: 98 Resp: 789292

Ion Ratio Lower Upper

98 100

100 63.9 51.7 77.5



#62

4-Bromofluorobenzene

Concen: 49.807 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN055461.D

Acq: 8 May 2019 2:00




Tgt Ion: 95 Resp: 244983


Ion Ratio Lower Upper

95 100

174 81.6 0.0 166.6

176 79.9 0.0 157.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN055461.D
Acq: 8 May 2019 2:00

Instrument :
MSVOA_N
ClientSampleId :
P026-SS002-0612-01

Tot Ion:117 Resp: 554921
Ion Ratio Lower Upper
117 100
82 58.4 46.2 69.4
119 32.1 26.6 40.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN055461.D
Acq: 8 May 2019 2:00

Tgt Ion:152 Resp: 187628
Ion Ratio Lower Upper
152 100
115 58.2 29.0 87.0
150 157.0 0.0 345.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

