


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN050719\
 Data File : VN055428.D
 Acq On : 7 May 2019 8:26
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Title : SW846 8260
 Last Update : Tue Apr 30 07:16:52 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	6617	PASS
75	95	30	60	52.5	16590	PASS
95	95	100	100	100.0	31618	PASS
96	95	5	9	6.8	2150	PASS
173	174	0.00	2	1.8	463	PASS
174	95	50	100	81.0	25610	PASS
175	174	5	9	7.9	2034	PASS
176	174	95	101	96.4	24688	PASS
177	176	5	9	6.4	1582	PASS