

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
 Data File : VN061330.D
 Acq On : 7 May 2020 13:47
 Operator : JC/MD
 Sample : PB128697ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128697ZHE#04

Quant Time: May 08 09:09:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	100513	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	186291	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	184197	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	76809	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	86681	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
35) Dibromofluoromethane	7.55	113	52854	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
50) Toluene-d8	10.06	98	239351	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
62) 4-Bromofluorobenzene	12.38	95	87562	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
Target Compounds						
16) Acetone	3.78	43	8107	7.828	ug/l	93
20) Methylene Chloride	4.51	84	2858	2.305	ug/l	94
43) Isopropyl Acetate	8.13	43	34029	9.379	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
Data File : VN061330.D
Acq On : 7 May 2020 13:47
Operator : JC/MD
Sample : PB128697ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128697ZHE#04

Quant Time: May 08 09:09:33 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

