

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
 Data File : VN061331.D
 Acq On : 7 May 2020 14:10
 Operator : JC/MD
 Sample : PB128697ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128697ZHE#05

Quant Time: May 08 09:10:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	98992	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	181653	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	179991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	73917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	84075	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
35) Dibromofluoromethane	7.55	113	51702	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
50) Toluene-d8	10.06	98	232850	53.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.36%	
62) 4-Bromofluorobenzene	12.38	95	85108	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
Target Compounds						
16) Acetone	3.78	43	8505	8.707	ug/l	99
20) Methylene Chloride	4.50	84	2757	2.257	ug/l	96
43) Isopropyl Acetate	8.13	43	33284	9.408	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050720\
Data File : VN061331.D
Acq On : 7 May 2020 14:10
Operator : JC/MD
Sample : PB128697ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128697ZHE#05

Quant Time: May 08 09:10:53 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

